

ADVANCED ULTRASOUND SOLUTIONS INC.—Technical Bulletin

Technical and Support Services

Bulletin #: INFO-SC01

REVISION DATE: 12-08-2016

DATE: December 8, 2016

TO: Hospital and Medical Center Administrators

Author(s): Christopher Froian

SUBJECT: Storage Stability of SONO Ultrasound and Disinfecting Wipes

PRODUCT: SONO Ultrasound Wipes and SONO Disinfecting Wipes.

EPA REGISTRATION NUMBER 6836-340-89018

PROBLEM: Storage temperature requirements not in MSDS

RESULT: This material can be safely stored in a warehouse or room, with temperatures from 40° to 100° F

Bulletin #: INFO-SC01

WARNING: THIS PAGE CONTAINS COPYRIGHTED MATERIALS THAT ARE CONFIDENTIAL AND/OR PROPRIETARY. ANY RELEASE OR DISTRIBUTION OF THIS MATERIAL WITHOUT PERMISSION, IS A VIOLATION OF THE LAW.

Copyright© 2012 Advanced Ultrasound Solutions, INC.